MycoFog™

MycoFog™ Quick Start guide (v.MF 1B)

Please consult full manual at https://mycofog.com

Day before use: Charge MycoFog™ Device with supplied cable.

> 8 hours

Day of use: Take out all cell cultures from incubator, set Incubator to **37 °C** or **45°C**, turn off gas control and remove water pan or turn off active humidification.

1. Open MycoFog™

Twist then lift top

2. Pour MycoFog™ Reagent into reagent reservoir of MycoFog™

MycoFog™ products are for Research Use Only, not for diagnostic or therapeutics use.

MycoFog™

3. Remove the wick holder, insert wick, replace holder and press firmly. Replace top of MycoFog™.

4. Place MycoFog™ in incubator, mid-chamber, wait 5 minutes.

5 minutes

5. Press Start

6. Close incubator door set timer.

180 Minutes (or **240** mins if using 37 °C). For incubators < 120 L chamber volume, extend 45°C protocol to 4 hours, extend 37°C protocol to 5 hours.

7. Remove MycoFog™, if necessary return temperature to 37°C. Your Incubator is now decontaminated and ready for use!